

# Detection of HSV in Skin and Genital Lesions: Comparison of Culture vs. AmpliVue HSV 1+2

Christian Renaud MD MSc FRCPC

Virology Laboratory

CHU Sainte-Justine, Montréal QC

November 19th 2014



# Disclosure

- For the current evaluation kits were supplied by Quidel Corporation [San Diego, CA, USA]
- No other financial support was received in order to carry out this evaluation


# Molecular detection of mucocutaneous HSV lesions

- Canadian and Québécois Guidelines on sexually transmitted infections mention superiority of molecular methods over culture due to higher sensitivity and excellent specificity
- Both culture and molecular methods are recommended as diagnostic tools for sexually transmitted infections
- Why go molecular?
  - Faster
  - Higher sensitivity
  - Easier to perform
  - Less hands on time
  - Easier sample transport
- Drawbacks?
  - Cost
  - Risk of contamination of CSF if performed simultaneously with high viral load lesions

# CHU Sainte-Justine

- CHU Sainte-Justine Virology Laboratory performs 8000 viral cultures/year including 3000 cultures for HSV
- 5 days/week HSV in-house PCR for CSF
- 2 ABI7500 thermocyclers used almost to full capacity
- Goals:
  - Faster turn around time
  - Reduced hands-on time
  - No contamination issues
  - No budget for new instrument

# Evaluation of AmpliVue HSV 1+2



- Molecular Isothermal Amplification
- No thermocycler required
- Closed system

# AmpliVue – Anytime, Anywhere

## Molecular Assays

### Assay Mechanism

- Multiplexing capabilities
- Asymmetrical amplification – generates single-stranded amplicons
- Detection probes bind to amplicons, the hybrid complex is captured and visualized on the test and control lines

### Assay Features

- Internal process control included to ensure DNA amplification occurs
- Lyophilized Master Mix packaged in individual amplification tubes for ease of use and extended shelf life



# Helicase Dependent Amplification

## Step 1:

The Helicase enzyme breaks the Hydrogen bonds between the nucleotide base pairs “unwinding” or “unzipping” the DNA at 64° C.

## Step 2 and 3:


“Unwinding” exposes the single strand DNA to the Biotin-Labeled Primer and Polymerase that will extend the copy.

## Step 4:

Once replication is complete, the Helicase performs the procedure again, leaving an original Template DNA and a Biotin-Labeled “copy.”


## Step 5:

The Capture Probe attaches to the Biotin-Labeled DNA copy creating a Biotin/Capture Probe (FITC, DIG complex).



# How AmpliVue cartridge works – universal strip

- ① 60 minutes of Iso-Thermal Amplification produces many Biotin/FITC Complexes in the amplification tube.
- ② With the cartridge closed, a buffer will carry the Biotin/Capture Probe Complexes across the Label Pad where Streptavidin coated Red beads adhere to the Biotin.
- ③ The Complex travels up the strip, being drawn by the Absorbent Pad with the Streptavidin coated Red beads to the Test Lines which contains Anti-FITC or Anti-DIG.
- ④ Test lines will “Capture” the Complex and Biotin/Streptavidin beads adhering to the Complex at the FITC-Labeled Probes and DIG-Labeled Probes.
- ⑤ This process will result in a Red Line appearing at the T2 and/or T1 (Test) line and indicate a Positive result.



Lines are clearly visible near the limit of detection




Replicate 1

Replicate 2


Replicate 3

# Internal control

- The internal control reflects if amplification and detection occurred.
- Generating the test line but not the control can be read as positive because the target material was amplified and detected.
  - This may occur because the sample itself has an overabundance of target DNA which creates competition for the reagents in the master mix.
- Thus the T2 line is well established and not enough reagent was available to amplify and develop a Control line.
- The internal control is a process control and indicates the process of HDA was successful.



# Fast and easy – results in 1 hour



- Lysis tube and Lysis heat block is not needed
- Amplification is only 45 minutes, not 1 hour

# AmpliVue HSV 1+2

HSV-1 Cutaneous Lesions				HSV-1 Mucocutaneous Lesions			
Combines Sites - Cutaneous Lesion (N=340)	ELVIS viral culture Reference Method			Combined Sites – Mucocutaneous Lesions (N=785)	ELVIS viral culture Reference Method		
	AmpliVue HSV 1+2	POS	NEG		Total	AmpliVue HSV 1+2 Assay	POS
POS	30	9	39	POS	149	22	171
NEG	0	301	301	NEG	8	606	614
Total	30	310	340	Total	157	628	785
		95% CI				95% CI	
Sensitivity	100%	88.60%	100%	Sensitivity	94.90%	90.30%	97.40%
Specificity	97.10%	94.60%	98.50%	Specificity	96.50%	94.80%	97.70%

\*One thousand three hundred fifty-five (1355) specimens from symptomatic male and female patients were tested. The swab specimens have been categorized as cutaneous (skin lesion, genital – penis), or mucocutaneous (anorectal, genital – vaginal/cervical, nares, ocular, oral lesion and urethral). Nineteen (19) tests were considered invalid and were removed from the performance analysis.

\*\*The reference ELVIS viral culture used in this study was unable to detect co-infected specimens. As a result, if a specimen was positive for HSV-2 it was removed from the calculation of HSV-1 clinical performance. Two hundred eleven (211) specimens identified as HSV-2 positive by ELVIS have been removed from the initial one thousand three hundred thirty-six (1336) specimens. The data below is for the remaining one thousand one hundred twenty-five (1125) specimens.



# AmpliVue HSV 1+2

HSV-2 Cutaneous Lesions				HSV-2 Mucocutaneous Lesions			
Combines Sites - Cutaneous Lesion (N=399)	ELVIS viral culture Reference Method			Combines Sites - Mucocutaneous Lesion (N=937)	ELVIS viral culture Reference Method		
	AmpliVue HSV 1+2	POS	NEG		Total	AmpliVue HSV 1+2	POS
POS	58	15	73	POS	148	33	181
NEG	1	325	326	NEG	4	752	756
Total	59	340	399	Total	152	785	937
		<b>95% CI</b>				<b>95% CI</b>	
Sensitivity	98.30%	91.00%	99.70%	Sensitivity	97.40%	93.40%	99.00%
Specificity	95.60%	92.80%	97.30%	Specificity	95.80%	94.20%	97.00%

\*One thousand three hundred fifty-five (1355) specimens from symptomatic male and female patients were tested. The swab specimens have been categorized as cutaneous (skin lesion, genital – penis), or mucocutaneous (anorectal, genital – vaginal/cervical, nares, ocular, oral lesion and urethral). Nineteen (19) tests were considered invalid and were removed from the performance analysis.

\*\*The reference ELVIS viral culture used in this study was unable to detect co-infected specimens. As a result, if a specimen was positive for HSV-2 it was removed from the calculation of HSV-1 clinical performance. Two hundred eleven (211) specimens identified as HSV-2 positive by ELVIS have been removed from the initial one thousand three hundred thirty-six (1336) specimens. The data below is for the remaining one thousand one hundred twenty-five (1125) specimens.

# Evaluation of AmpliVue HSV 1+2 in comparison to viral culture

- Testing of 323 consecutive mucocutaneous samples with viral culture and AmpliVue (reflex testing performed with an in-house PCR)
- Samples were received in 3 mL UTM
- Viral Culture set up (everyday of the week: 7/7)
  - VERO cell in-house
  - MRHF commercial (Diagnostic Hybrids/Quidel Corporation)
  - Kept for 5 days
  - DFA confirmation (PathoDx) via CPE
- Aliquot for Amplivue (4°C) performed 5/7
- Aliquot for in-house PCR (-20°C) and batch testing

# Sample Types

113	lesions (lips, buttocks, labia, thigh, cutaneous, scrotum, foreskin...)
91	vulva
27	penis
25	throat/mouth
24	vagina
12	ano-rectal
10	unspecified
6	genital (unspecified)
5	eye/conjunctiva
4	tongue
4	cervix
2	urethra

# HSV-1 Calculations

<b>Sensitivity</b>	$\frac{a}{a + c}$	= <b>98.65 %</b>	95% CI: <b>92.67 %</b> to <b>99.77 %</b>
<b>Specificity</b>	$\frac{d}{b + d}$	= <b>96.67 %</b>	95% CI: <b>93.54 %</b> to <b>98.55 %</b>
<b>Positive Likelihood Ratio</b>	$\frac{\text{Sensitivity}}{100 - \text{Specificity}}$	= <b>29.59</b>	95% CI: <b>14.97</b> to <b>58.52</b>
<b>Negative Likelihood Ratio</b>	$\frac{100 - \text{Sensitivity}}{\text{Specificity}}$	= <b>0.01</b>	95% CI: <b>0.00</b> to <b>0.10</b>
<b>Disease prevalence</b>	$\frac{a + c}{a + b + c + d}$	= <b>23.57 % (*)</b>	95% CI: <b>18.98 %</b> to <b>28.66 %</b>
<b>Positive Predictive Value</b>	$\frac{a}{a + b}$	= <b>90.12 % (*)</b>	95% CI: <b>81.46 %</b> to <b>95.63 %</b>
<b>Negative Predictive Value</b>	$\frac{d}{c + d}$	= <b>99.57 % (*)</b>	95% CI: <b>97.62 %</b> to <b>99.93 %</b>

\*9 total invalids (2.8%): All calculations excluding invalid results. Invalid results most likely due to user processing error or assay inhibition.

# HSV-1

		Culture	
		Pos	Neg
Amplivue	Pos	73	8
	Neg/invalid	3	239*

9 invalid (2.8%)\*

**Kappa= 0.923** SE of kappa = 0.025; 95% confidence interval: From 0.873 to 0.972 The strength of agreement is considered to be « very good. »

OPA (overall percentage agreement) = 97.1

PPA (positive percentage agreement) = 98.6

NPA (negative percentage agreement)=96.6

\*9 total invalids (2.8%): All calculations excluding invalid results. Invalid results most likely due to user processing error or assay inhibition.

# HSV-2 Calculations

<b>Sensitivity</b>	$\frac{a}{a + c}$	= <b>100.00 %</b>	95% CI: <b>90.17 % to 100.00 %</b>
<b>Specificity</b>	$\frac{d}{b + d}$	= <b>98.20 %</b>	95% CI: <b>95.85 % to 99.41 %</b>
<b>Positive Likelihood Ratio</b>	$\frac{\text{Sensitivity}}{100 - \text{Specificity}}$	= <b>55.60</b>	95% CI: <b>23.33 to 132.53</b>
<b>Negative Likelihood Ratio</b>	$\frac{100 - \text{Sensitivity}}{\text{Specificity}}$	= <b>0.00</b>	
<b>Disease prevalence</b>	$\frac{a + c}{a + b + c + d}$	= <b>11.46 % (*)</b>	95% CI: <b>8.16 % to 15.52 %</b>
<b>Positive Predictive Value</b>	$\frac{a}{a + b}$	= <b>87.80 % (*)</b>	95% CI: <b>73.78 % to 95.87 %</b>
<b>Negative Predictive Value</b>	$\frac{d}{c + d}$	= <b>100.00 % (*)</b>	95% CI: <b>98.64 % to 100.00 %</b>

\*9 total invalids (2.8%): All calculations excluding invalid results. Invalid results most likely due to user processing error or assay inhibition.

# HSV-2

		Culture	
		Pos	Neg
AmpliVue	Pos	36	5
	Neg/Invalid	0	282

9 invalids (2.8%)\*

**Kappa= 0.926** SE of kappa = 0.033; 95% confidence interval: from 0.862 to 0.990. The strength of agreement is considered to be « very good. »

OPA (overall percentage agreement) = 98.4

PPA (positive percentage agreement) = 100

NPA (negative percentage agreement) = 98.3

\*9 total invalids (2.8%): All calculations excluding invalid results. Invalid results most likely due to user processing error or assay inhibition.

# HSV 1+2 Calculations

<b>Sensitivity</b>	$\frac{a}{a + c}$	= <b>99.09 %</b>	95% CI: <b>95.02 % to 99.85 %</b>
<b>Specificity</b>	$\frac{d}{b + d}$	= <b>93.63 %</b>	95% CI: <b>89.35 % to 96.56 %</b>
<b>Positive Likelihood Ratio</b>	$\frac{\text{Sensitivity}}{100 - \text{Specificity}}$	= <b>15.55</b>	95% CI: <b>9.19 to 26.32</b>
<b>Negative Likelihood Ratio</b>	$\frac{100 - \text{Sensitivity}}{\text{Specificity}}$	= <b>0.01</b>	95% CI: <b>0.00 to 0.07</b>
<b>Disease prevalence</b>	$\frac{a + c}{a + b + c + d}$	= <b>35.03 % (*)</b>	95% CI: <b>29.76 % to 40.59 %</b>
<b>Positive Predictive Value</b>	$\frac{a}{a + b}$	= <b>89.34 % (*)</b>	95% CI: <b>82.46 % to 94.20 %</b>
<b>Negative Predictive Value</b>	$\frac{d}{c + d}$	= <b>99.48 % (*)</b>	95% CI: <b>97.12 % to 99.91 %</b>

\*9 total invalids (2.8%): All calculations excluding invalid results. Invalid results most likely due to user processing error or assay inhibition.

# HSV 1 + 2

		Culture	
		Pos	Neg
AmpliVue	Pos	109	13
	Neg/Invalid	3	198

9 invalids (2.8%)\*

Potentially 4% more positives

**Kappa= 0.904** SE of kappa = 0.025 ; 95% confidence interval: From 0.856 to 0.953 The strength of agreement is considered to be « very good. »

OPA (overall percentage agreement) = 95.5

PPA (positive percentage agreement) = 99.1

NPA (negative percentage agreement) = 93.6

\*9 total invalids (2.8%): All calculations excluding invalid results. Invalid results most likely due to user processing error or assay inhibition.

# Conclusion

- Amplivue HSV 1+2 has « very good » agreement with viral culture (*VERO+MRHF for 5 days*) for each target separately and both combined targets
- Sensitivity/Specificity is :
  - 96%/96% for HSV-1
  - 100%/98.2% for HSV-2
  - 97.3%/93.8% for HSV 1+2
- 2.8% of invalid results (not any specific sample types, mostlikely processing errors)
- Comparable sensitivity is expected; potentially 4% more positives
- Very easy to perform (batch twice a day)